

**ATTACHMENT II  
TO APPENDIX B OF THE UNOS BYLAWS**

**Criteria for Designated Histocompatibility Laboratories**

A histocompatibility laboratory that meets the following criteria shall be qualified as a designated histocompatibility laboratory to perform histocompatibility testing for designated transplant programs.

- III. Facilities and Resources.** The size and training of the staff must be sufficient to perform the volume and variety of tests required without a degree of pressure that will result in errors. Laboratory space must be sufficient so that all procedures can be carried out without crowding to the extent that errors may result.

The laboratory must establish and employ policies and procedures for the proper maintenance of equipment, instruments and test systems by 1) defining its preventive maintenance program for each instrument and piece of equipment, and by 2) performing and documenting function checks on equipment with at least the frequency specified by the manufacturer.